

QUALITY CONTROL CERTIFICATE

FINAL PRODUCT DATA

Description	PLATE COUNT AGAR (STANDARD METHODS AGAR) (ISO 4833)
Code	010088
Lot number	2791k128
Production date	28/03/24
Expiration date	28/07/24

PHYSICAL AND CHEMICAL CONTROL OF FINAL PRODUCT

Control	Expected results	Observed results
Final pH at 25°C	7.0 +/- 0,2	7.0 (Confirmed)
Quantity of material	20ml (010088)	Confirmed
Color & cleanliness of the material	Beige open clear.	Confirmed

STERILITY TEST OF FINAL PRODUCTS

Number of controlled products (3%)	Incubation	Number of suitable products	Number of inappropriate products
4/010088	36 ± 1 °C for 48 hours	4/010088	0

FINAL PRODUCT PERFORMANCE TEST (According to ISO 4833) Incubate: 30 ± 1 °C for 72 hours

Microorganism	Expected results	Observed results
<i>Escherichia coli</i> (≥500 cfu/10μl) e10.24 ATCC 25922. Lot: 00207501. (Productivity)	PR ≥ 0,7	426 cfu
<i>Staphylococcus aureus</i> (≥500 cfu/10μl) e10.24 ATCC: 29213. Lot: 01403703 (productivity)	PR ≥ 0,7	383 cfu

- Production and quality control tests were carried out at the premises of Bioprep G. Papanikolaou & Co. According to EN ISO 9001: 2008. In addition, the company's audit department implements to ISO 11133: 2014.
- The products of this batch will have the results reported once the user receives them cool (8-25 °C) then keep them at the temperature indicated on the packaging until the expiration date and follow the using instructions of the brochure.
- PR: Productivity Rate. IV: Inoculum Value
- 0: Growth zero. 1: Weak growth. 2: Good growth. PR: Productivity Rate.


Release of the batch

YES NO

DATE

01/04/24

Iniotaki Fofi



Quality Control Manager