

**QUALITY CONTROL CERTIFICATE****FINAL PRODUCT DATA**

Description	TRYPTONE SOY AGAR (USP/EP)
Code	010119
Lot number	2769m123
Production date	23/02/24
Expiration date	23/06/24

**PHYSICAL AND CHEMICAL CONTROL OF FINAL PRODUCT**

Control	Expected results	Observed results
Final pH at 25°C	7.3 +/- 0,2	7.2 (Confirmed)
Quantity of material	20ml (010119)	Confirmed
Color & cleanliness of the material	Beige clear	Confirmed

**STERILITY TEST OF FINAL PRODUCTS**

Number of controlled products (‰)	Incubation	Number of suitable products	Number of inappropriate products
9/010119	36 ± 1 °C for 48 hours	5/010119	0


**FINAL PRODUCT PERFORMANCE TEST Incubate: (30-35) °C for (18-24) hours**

Microorganism	Expected results	Observed results
<i>Escherichia coli</i> (≥500 cfu/10μl) e10.24 ATCC: 25922. Lot: 00207501 (productivity)	PR ≥ 0,7	429 cfu
<i>Staphylococcus aureus</i> (≥500 cfu/10μl) e10.24 ATCC: 29213. Lot: 01403703 (productivity)	PR ≥ 0,7	396 cfu

- Production and quality control tests were carried out at the premises of Bioprep G. Papanikolaou & Co. According to EN ISO 9001:2015 / ΕΛΟΤ EN ISO 13485:2016. In addition, the company's audit department implements to ISO 11133: 2014.
- The products of this batch will have the results reported once the user receives them cool (8-25 °C) then keep them at the temperature indicated on the packaging until the expiration date and follow the using instructions of the brochure.
- PR: Productivity Rate. IV: Inoculum Value

**Release of the batch**YES  NO **DATE**

27/02/2024

**Iniotaki Fofi****Quality Control Manager**