

**QUALITY CONTROL CERTIFICATE****FINAL PRODUCT DATA**

Description	RPMI GLUCOSE AGAR
Code	010293
Lot number	2795d103
Production date	03/04/24
Expiration date	03/07/24

**PHYSICAL AND CHEMICAL CONTROL OF FINAL PRODUCT**

Control	Expected results	Observed results
Final pH at 25°C	7,0 +/- 0,2	6.9 (Confirmed)
Quantity of material	20ml (010293)	Confirmed
Color & cleanliness of the material	Light pink clear	Confirmed

**STERILITY TEST OF FINAL PRODUCTS**

Number of controlled products (3‰)	Incubation	Number of suitable products	Number of inappropriate products
4/010293	36 ± 1 °C for 48 hours	4/010293	0

**FINAL PRODUCT PERFORMANCE TEST Incubate at 35 - 37 °C for 48 h. ETest for 48h. MIC limits to (µg/ml).**

Microorganism	Expected results		Observed results	
	Flucytosine 0,002 - 32 µg/ml	Amphotericin B 0.002 - 32 µg/mL	Flucytosine 0,002 - 32 µg/ml	Amphotericin B 0.002 - 32 µg/mL
<i>Candida albicans</i> ATCC: 10231. Lot: 04205401 e10.24 (Θολερότητα RPMI 0,5 McFarlant)	0,5<CMI - MIC<2	0,125<CMI - MIC<0,5	0,27=CMI – MIC=1,1 (Conform)	0,129=CMI - MIC=0,5 (Conform)
<i>Candida parapsilosis</i> ATCC® 22019. Lot: 02505203 e10.24 (Θολερότητα RPMI 0,5 McFarlant)	0,064<CMI - MIC<0,25	0,250<CMI - MIC<1	0,042=CMI - MIC=0,16 (Conform)	0,130=CMI - MIC=0,76 (Conform)

- Production and quality control tests were carried out at the premises of Bioprep G. Papanikolaou & Co. According to EN ISO 9001: 2008. In addition, the company's audit department implements to ISO 11133: 2014.
- The products of this batch will have the results reported once the user receives them cool (8-25 °C) then keep them at the temperature indicated on the packaging until the expiration date and follow the using instructions of the brochure.
- 0: Growth zero. 1: Weak growth. 2: Good growth. PR: Productivity Rate.


Release of the batch

YES  NO 

DATE

05/04/2024

Iniotaki Fofi



Quality Control Manager