

QUALITY CONTROL CERTIFICATE**FINAL PRODUCT DATA**

Description	CHROMagar™ MRSA
Code	010346
Lot number	2751n230
Production date	30/01/24
Expiration date	30/03/24

PHYSICAL AND CHEMICAL CONTROL OF FINAL PRODUCT

Control	Expected results	Observed results
Final pH at 25°C	7,0 +/- 0,2	7.0 (Confirmed)
Quantity of material	20ml (010346)	Confirmed
Color & cleanliness of the material	Agar straw chrome clear	Confirmed

STERILITY TEST OF FINAL PRODUCTS

Number of controlled products (3‰)	Incubation	Number of suitable products	Number of inappropriate products
2/010346	36 ± 1 °C for 48 hours	2/010346	0

FINAL PRODUCT PERFORMANCE TEST Incubate at 36 ± 1 °C for 18-48 h

Microorganism	Expected results	Observed results
<i>Staphylococcus aureus</i> (MRSA) (≥500 cfu/10μl) ATCC: 12493. Lot: 06401102 (productivity)	PR ≥ 0,5 Pink to mauve Colony Color	Pink to mauve Colony Color 268 cfu
<i>Staphylococcus aureus</i> (≥500 cfu/10μl) ATCC: 29213. Lot: 01403703 (selectivity)	Inhibited (0)	Confirmed 0 cfu

- Production and quality control tests were carried out at the premises of Bioprep G. Papanikolaou & Co. According to EN ISO 9001:2015 / ΕΛΟΤ EN ISO 13485:2016. In addition, the company's audit department implements to ISO 11133: 2014.
- The products of this batch will have the results reported once the user receives them cool (8-25 °C) then keep them at the temperature indicated on the packaging until the expiration date and follow the using instructions of the brochure.
- 0: Growth zero. 1: Weak growth. 2: Good growth. PR: Productivity Rate.


Release of the batch

YES NO

DATE

04/02/2024

Iniotaki Fofi



Quality Control Manager