

**QUALITY CONTROL CERTIFICATE****FINAL PRODUCT DATA**

Description	CHROMagar™ Strep B
Code	010362
Lot number	2750c129
Production date	29/01/24
Expiration date	29/03/24

**PHYSICAL AND CHEMICAL CONTROL OF FINAL PRODUCT**

Control	Expected results	Observed results
Final pH at 25°C	7,3 +/- 0,2	7,3 (Confirmed)
Quantity of material	20ml (010362)	Confirmed
Color & cleanliness of the material	Agar gray clear	Confirmed

**STERILITY TEST OF FINAL PRODUCTS**

Number of controlled products (3‰)	Incubation	Number of suitable products	Number of inappropriate products
2/010362	36 ± 1 °C for 48 hours	2/010362	0

**FINAL PRODUCT PERFORMANCE TEST. Incubate: (37 ± 1) °C for (24 ± 2) h**

Microorganism	Expected results	Observed results
<i>Streptococcus agalactiae</i> (≥500 cfu/10μl) ATCC: 13813. Lot: 01603402,b (productivity)	PR ≥ 0,5 Pink – mauve Colony Color	Pink – mauve Colony Color 291 cfu
<i>Escherichia coli</i> (≥500 cfu/10μl) ATCC: 25922. Lot: 00207501 (Selectivity)	Inhibited (0)	Confirmed 0 cfu

- Production and quality control tests were carried out at the premises of Bioprep G. Papanikolaou & Co. According to EN ISO 9001:2015 / ΕΛΟΤ EN ISO 13485:2016. In addition, the company's audit department implements to ISO 11133: 2014.
- The products of this batch will have the results reported once the user receives them cool (8-25 °C) then keep them at the temperature indicated on the packaging until the expiration date and follow the using instructions of the brochure.
- 0: Growth zero. 1: Weak growth. 2: Good growth. PR: Productivity Rate.


Release of the batch

YES  NO 

DATE

03/02/2024

Iniotaki Fofi



Quality Control Manager