

**QUALITY CONTROL CERTIFICATE****FINAL PRODUCT DATA**

Description	SABOURAUD DEXTROSE AGAR W/ CHLORAMPHENICOL
Code	050098
Lot number	2768t122
Production date	22/02/24
Expiration date	22/06/24

**PHYSICAL AND CHEMICAL CONTROL OF FINAL PRODUCT**

Control	Expected results	Observed results
Final pH at 25°C	5,6 +/- 0,2	5,6 (Confirmed)
Quantity of material	10ml (050098)	Confirmed
Color & cleanliness of the material	Clear beige agar.	Confirmed

**STERILITY TEST OF FINAL PRODUCTS**

Number of controlled products (3‰)	Incubation	Number of suitable products	Number of inappropriate products
2/050039	36 ± 1 °C for 48 hours	2/050039	0


**FINAL PRODUCT PERFORMANCE TEST** Incubate: 36 ± 1 °C for 24 - 48 h

Microorganism	Expected results	Observed results
<i>Candida albicans</i> (≥500 cfu/10μl) e10.24 ATCC: 10231. Lot: 04205401 (productivity)	PR ≥ 0,7 Typical colonies consistent with the species	Typical colonies consistent with the species 386 cfu
<i>Escherichia coli</i> (≥500 cfu/10μl) e10.24 ATCC 25922. Lot: 00207501 (selectivity)	Inhibited (0)	0 cfu

- Production and quality control tests were carried out at the premises of Bioprep G. Papanikolaou & Co. According to EN ISO 9001:2015 / ΕΛΟΤ EN ISO 13485:2016. In addition, the company's audit department implements to ISO 11133: 2014.
- The products of this batch will have the results reported once the user receives them cool (8-25 °C) then keep them at the temperature indicated on the packaging until the expiration date and follow the using instructions of the brochure.
- 0: Growth zero. 1: Weak growth. 2: Good growth. PR: Productivity Rate.

**Release of the batch**YES  NO **DATE**

27/02/2024

**Iniotaki Fofi****Quality Control Manager**