

QUALITY CONTROL CERTIFICATE**FINAL PRODUCT DATA**

Description	SABOURAUD DEXTROSE AGAR (USP/EP)
Code	060292
Lot number	2771d127
Production date	27/02/24
Expiration date	27/02/25

PHYSICAL AND CHEMICAL CONTROL OF FINAL PRODUCT

Control	Expected results	Observed results
Final pH at 25°C	5,6 +/- 0,2	5,5 (Confirmed)
Quantity of material	100ml (060292)	Confirmed
Color & cleanliness of the material	Beige clear	Confirmed

STERILITY TEST OF FINAL PRODUCTS

Number of controlled products (3‰)	Incubation	Number of suitable products	Number of inappropriate products
1/060292	36 ± 1 °C for 48 hours	1/060292	0

FINAL PRODUCT PERFORMANCE TEST

Microorganism	Expected results	Observed results
<i>Candida albicans</i> (≥500 cfu/10μl) e10.24 ATCC: 10231. Lot: 04205401 (productivity) Επίσπαση: 35-37 °C για 24-48 ώρες.	PR ≥ 0,7	395 cfu
<i>Escherichia coli</i> (≥500 cfu/10μl) e10.24 ATCC 25922. Lot: 00207501 (selectivity) Επίσπαση: 35-37 °C για 24-48 ώρες.	PR ≥ 0,1	65 cfu

- Production and quality control tests were carried out at the premises of Bioprep G. Papanikolaou & Co. According to EN ISO 9001:2015 / ΕΛΟΤ EN ISO 13485:2016. In addition, the company's audit department implements to ISO 11133: 2014.
- The products of this batch will have the results reported once the user receives them cool (8-25 °C) then keep them at the temperature indicated on the packaging until the expiration date and follow the using instructions of the brochure.
- 0: Growth zero. 1: Weak growth. 2: Good growth


Release of the batch

YES NO

DATE

02/03/24

Iniotaki Fofi



Quality Control Manager