

**QUALITY CONTROL CERTIFICATE****FINAL PRODUCT DATA**

Description	TRICHOMONAS BROTH
Code	080117
Lot number	2789r226
Production date	26/03/24
Expiration date	26/07/24

**PHYSICAL AND CHEMICAL CONTROL OF FINAL PRODUCT**

Control	Expected results	Observed results
Final pH at 25°C	7.0 +/- 0,2	7.0 (Confirmed)
Quantity of material	3ml (080117)	Confirmed
Color & cleanliness of the material	Beige clear with pink ring	Confirmed

**STERILITY TEST OF FINAL PRODUCTS**

Number of controlled products (3%)	Incubation	Number of suitable products	Number of inappropriate products
2/080117	36 ± 1 °C for 48 hours	2/080117	0


**FINAL PRODUCT PERFORMANCE TEST (Incubate at 32 - 36 °C for 1 - 4 days)**

Microorganism	Expected results	Observed results
<i>Trichomonas vaginalis</i> (≤10 cfu/ml) ATCC 30001 (productivity)	Growing (1-2) (Moderate to good)	Confirmed (Growth in TB checked)
<i>Escherichia coli</i> (≤1000 cfu/ml) ATCC: 25922. Lot: 00207501	Inhibited (0)	Confirmed (Growth in TSB checked)

- Production and quality control tests were carried out at the premises of Bioprep G. Papanikolaou & Co. According to EN ISO 9001:2015 / ΕΛΟΤ EN ISO 13485:2016. In addition, the company's audit department implements to ISO 11133: 2014.
- The products of this batch will have the results reported once the user receives them cool (8-25 °C) then keep them at the temperature indicated on the packaging until the expiration date and follow the using instructions of the brochure.
- 0: Growth zero. 1: Weak growth. 2: Good growth

**Release of the batch**YES  NO **DATE**

30/03/2024

**Iniotaki Fofi****Quality Control Manager**