

QUALITY CONTROL CERTIFICATE**FINAL PRODUCT DATA**

Description	TRYPTONE SOY AGAR (USP/EP)
Code	010119
Lot number	2740c115
Production date	15/01/24
Expiration date	15/05/24

PHYSICAL AND CHEMICAL CONTROL OF FINAL PRODUCT

Control	Expected results	Observed results
Final pH at 25°C	7.3 +/- 0,2	7.1 (Confirmed)
Quantity of material	20ml (010119)	Confirmed
Color & cleanliness of the material	Beige clear	Confirmed

STERILITY TEST OF FINAL PRODUCTS

Number of controlled products (‰)	Incubation	Number of suitable products	Number of inappropriate products
4/010119	36 ± 1 °C for 48 hours	4/010119	0


FINAL PRODUCT PERFORMANCE TEST Incubate: (30-35) °C for (18-24) hours

Microorganism	Expected results	Observed results
<i>Escherichia coli</i> (≥500 cfu) ATCC: 25922. Lot: 00207501 (productivity)	PR ≥ 0,7	412 cfu
<i>Staphylococcus aureus</i> (≥500 cfu) ATCC: 29213. Lot: 01403703 (productivity)	PR ≥ 0,7	391 cfu

- Production and quality control tests were carried out at the premises of Bioprep G. Papanikolaou & Co. According to EN ISO 9001:2015 / ΕΛΟΤ EN ISO 13485:2016. In addition, the company's audit department implements to ISO 11133: 2014.
- The products of this batch will have the results reported once the user receives them cool (8-25 °C) then keep them at the temperature indicated on the packaging until the expiration date and follow the using instructions of the brochure.
- PR: Productivity Rate. IV: Inoculum Value

Release of the batchYES NO **DATE**

20/01/24

Iniotaki Fofi**Quality Control Manager**